



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

CALIBRATION LABORATORY, GLASSCO LABORATORY EQUIPMENTS PRIVATE LIMITED, MANGLAI, P. O. KHUDDA KALAN, AMBALA CANTT, HARYANA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

NABLC0626HR04802

Page No

1 of 5

Validity

24/07/2026 to 23/07/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
Permanent Facility					
1	MECHANICAL-VOLUME	Density Bottle / Pycnometer	Using Digital Weighing Balance up to 220 g readability 0.1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	10 ml to 100 ml	0.003 ml
2	MECHANICAL-VOLUME	Glass Burette	Using Digital Weighing Balance up to 220 g readability 0.1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	10 ml to 100 ml	0.003 ml
3	MECHANICAL-VOLUME	Graduated Pipettes	Using Digital Weighing Balance up to 220 g readability 0.1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	>10 ml to 50 ml	0.002 ml



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

CALIBRATION LABORATORY, GLASSCO LABORATORY EQUIPMENTS PRIVATE LIMITED, MANGLAI, P. O. KHUDDA KALAN, AMBALA CANTT, HARYANA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

NABLC0626HR04802

Page No

2 of 5

Validity

24/07/2026 to 23/07/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
4	MECHANICAL-VOLUME	Graduated Pipettes	Using Digital Weighing Balance up to 220 g readability 0.1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	0.1 ml to 10 ml	0.001 ml
5	MECHANICAL-VOLUME	Volumetric Flasks / Measuring Cylinders	Using Digital Weighing Balance up to 220 g readability 0.1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	1 ml to 10 ml	0.001 ml
6	MECHANICAL-VOLUME	Volumetric Flasks/Measuring Cylinders	Using Digital Weighing Balance up to 220 g readability 0.1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	>10 ml to 100 ml	0.003 ml



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

CALIBRATION LABORATORY, GLASSCO LABORATORY EQUIPMENTS PRIVATE LIMITED, MANGLAI, P. O. KHUDDA KALAN, AMBALA CANTT, HARYANA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

NABLC0626HR04802

Page No

3 of 5

Validity

24/07/2026 to 23/07/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
7	MECHANICAL-VOLUME	Volumetric Flasks/Measuring Cylinders	Using Digital Weighing Balance up to 1020 g readability 1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	>100 ml to 250 ml	0.006 ml
8	MECHANICAL-VOLUME	Volumetric Flasks/Measuring Cylinders	Using Digital Weighing Balance up to 4200 g and readability 10 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	>1000 ml to 2000 ml	0.068 ml
9	MECHANICAL-VOLUME	Volumetric Flasks/Measuring Cylinders	Using Digital Weighing Balance up to 8100 g readability 10 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	>2000 ml to 6000 ml	0.15 ml



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

CALIBRATION LABORATORY, GLASSCO LABORATORY EQUIPMENTS PRIVATE LIMITED, MANGLAI, P. O. KHUDDA KALAN, AMBALA CANTT, HARYANA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

NABLC0626HR04802

Page No

4 of 5

Validity

24/07/2026 to 23/07/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
10	MECHANICAL-VOLUME	Volumetric Flasks/Measuring Cylinders	Using Digital Weighing Balance up to 1020 g readability 1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	>250 ml to 500 ml	0.017 ml
11	MECHANICAL-VOLUME	Volumetric Flasks/Measuring Cylinders	Using Digital Weighing Balance up to 2100 g readability 1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	>500 ml to 1000 ml	0.03 ml
12	MECHANICAL-VOLUME	Volumetric Pipette	Using Digital Weighing Balance up to 220 g readability 0.1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	>10 ml to 100 ml	0.003 ml



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : CALIBRATION LABORATORY, GLASSCO LABORATORY EQUIPMENTS PRIVATE LIMITED, MANGLAI, P. O. KHUDDA KALAN, AMBALA CANTT, HARYANA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number NABLC0626HR04802

Validity 24/07/2026 to 23/07/2030

Page No 5 of 5

Last Amended on -

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
13	MECHANICAL-VOLUME	Volumetric Pipette	Using Digital Weighing Balance up to 420 g readability 1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	>100 ml to 200 ml	0.006 ml
14	MECHANICAL-VOLUME	Volumetric Pipette	Using Digital Weighing Balance up to 220 g readability 0.1 mg and distilled water of known density as per ISO 4787:2021 & ISO/TR 20461:2023	0.5 ml to 10 ml	0.002 ml

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of $k = 2$.